

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/26/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:55
INTERNAL STANDARD	VERIFIED	02:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:56
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:57
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:58
BLANK TEST	0.00	02:58



605 *Ralph Bartlett* WESO #162 12-24-10