

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 01/04/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:05
INTERNAL STANDARD	VERIFIED	02:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:07
BLANK TEST	0.00	02:08
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:08
BLANK TEST	0.00	02:09



SGS Resp of Results 01-04-11

WCSO #162