

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 01/18/2011

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:19
INTERNAL STANDARD	VERIFIED	02:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:20
BLANK TEST	0.00	02:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:21
BLANK TEST	0.00	02:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:22



SOB High Brass

1/18/11

WASO #162