

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/07/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10270

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:03
INTERNAL STANDARD	VERIFIED	02:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:04
BLANK TEST	0.00	02:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:05
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07



565 *Ralph Bartlett*

WCSO #162

02-07-11