

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/21/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10270

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:07
INTERNAL STANDARD	VERIFIED	02:07
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:09
BLANK TEST	0.00	02:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:10
BLANK TEST	0.00	02:11



SGT Rugh / Bunter WCSO #142 02/21/11