

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 03/15/2011

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 10270

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:30
INTERNAL STANDARD	VERIFIED	03:30
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	03:30
BLANK TEST	0.00	03:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:31
BLANK TEST	0.00	03:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:33



005 Rust J Bartlett

03-15-11

WCSO #112