

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/21/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10270

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:10
INTERNAL STANDARD	VERIFIED	02:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:11
BLANK TEST	0.00	02:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:12
BLANK TEST	0.00	02:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:13
BLANK TEST	0.00	02:14



SGS Ralph J Bartlett WCSO SHERIFF #162 3-21-11