

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/07/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:39
INTERNAL STANDARD	VERIFIED	02:39
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:42
BLANK TEST	0.00	02:42



SOS *Ralph Bartlett*

WCSO SERIAL # 162

4/7/11