

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/13/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:53
INTERNAL STANDARD	VERIFIED	01:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:53
BLANK TEST	0.00	01:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:54
BLANK TEST	0.00	01:55
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:55
BLANK TEST	0.00	01:56



Sgt. Ralph Bartlett WCSO #162

04-13-11