

SUPERVISOR TEST

WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 105208

Date: 04/27/2011

Operator Name: JOSHUA M LOPEZ  
 Permit #: 34728  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:49
INTERNAL STANDARD	VERIFIED	04:49
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:49
BLANK TEST	0.00	04:50
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	04:51
BLANK TEST	0.00	04:51
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	04:52
BLANK TEST	0.00	04:52

*J. M. Lopez #295*

