

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 05/02/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:33
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	03:34
BLANK TEST	0.00	03:35



SGS *Ralph Bartlett* WOSO SHERIFF #102 5/2/11