

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/11/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:56
INTERNAL STANDARD	VERIFIED	02:56
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:58
BLANK TEST	0.00	02:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:59
BLANK TEST	0.00	02:59



SOB RALPH BARTLETT WOSO #162 5-11-11