

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 05/16/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:49
INTERNAL STANDARD	VERIFIED	01:49
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:50
REFERENCE STANDARD TEMP	33.93c	
REFERENCE STANDARD	0.09	01:51
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	33.93c	
REFERENCE STANDARD	0.09	01:52
BLANK TEST	0.00	01:52



SGT Ralph J Bartlett WESO #162 05/16/11