

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/08/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:19
INTERNAL STANDARD	VERIFIED	02:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:19
BLANK TEST	0.00	02:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:21
BLANK TEST	0.00	02:21
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:22



SGS Ralph Bartlett WCSI #162 06-08-11