

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/22/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:03
INTERNAL STANDARD	VERIFIED	02:03
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:03
BLANK TEST	0.00	02:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:05
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:06



SOS Ralph Bartlett *WERO #100* *06/22/11*