

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 07/14/2011

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:37
INTERNAL STANDARD	VERIFIED	01:37
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	01:37
BLANK TEST	0.00	01:38
REFERENCE STANDARD TEMP	33.93c	
REFERENCE STANDARD	0.09	01:38
BLANK TEST	0.00	01:39
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:39
BLANK TEST	0.00	01:40



505 *Rough Print* use #162

07-14-11