

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/19/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:46
INTERNAL STANDARD	VERIFIED	02:46
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:48
BLANK TEST	0.00	02:48
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:49
BLANK TEST	0.00	02:50



SG *Ralph Bartlett* WCSO SHERIFF #162 07-19-11