

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 07/25/2011

Operator Name: JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:37
INTERNAL STANDARD	VERIFIED	02:37
REFERENCE STANDARD TEMP	33.93c	
REFERENCE STANDARD	0.09	02:38
BLANK TEST	0.00	02:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:40

*J. Lopez #299*

