

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 08/02/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:14
INTERNAL STANDARD	VERIFIED	00:14
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	00:14
BLANK TEST	0.00	00:15
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:15
BLANK TEST	0.00	00:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:16
BLANK TEST	0.00	00:17



Sgt. R. J. Bartlett

WER #162

08-02-11