

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/11/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:58
INTERNAL STANDARD	VERIFIED	00:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:58
BLANK TEST	0.00	00:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:59
BLANK TEST	0.00	01:00
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:00
BLANK TEST	0.00	01:01



SGT Ralph J Bartlett

WCSO SHERIFF #142

08-11-11