

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 08/15/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:59
INTERNAL STANDARD	VERIFIED	23:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:59
BLANK TEST	0.00	00:00
REFERENCE STANDARD TEMP	33.93c	
REFERENCE STANDARD	0.09	00:01
BLANK TEST	0.00	00:02
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:02
BLANK TEST	0.00	00:03



SUS Ralph J Bartlett 08-16-11

W010 #162