

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/22/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:49
INTERNAL STANDARD	VERIFIED	03:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:49
BLANK TEST	0.00	03:50
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	03:51
BLANK TEST	0.00	03:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:52
BLANK TEST	0.00	03:52

