

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/29/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:37
INTERNAL STANDARD	VERIFIED	23:37
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	23:37
BLANK TEST	0.00	23:38
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:38
BLANK TEST	0.00	23:39
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:39
BLANK TEST	0.00	23:40



SGC Ralph Bartlett WOSO STAAD #162 08/29/11