

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/08/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:54
INTERNAL STANDARD	VERIFIED	02:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:57



S65 Ralph J Bartlett WSO #162 09-08-11