

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/10/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:01
INTERNAL STANDARD	VERIFIED	22:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:02
BLANK TEST	0.00	22:02
REFERENCE STANDARD TEMP	33.93c	
REFERENCE STANDARD	0.09	22:03
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	22:04
BLANK TEST	0.00	22:04



CIT Russ Bartlett

WC10 #162

10/10/11