

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/20/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:04
INTERNAL STANDARD	VERIFIED	01:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:04
BLANK TEST	0.00	01:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:06
BLANK TEST	0.00	01:06
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:07



[Handwritten signature]

10/20/11