

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/25/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11140

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 22:08 |
| INTERNAL STANDARD       | VERIFIED | 22:08 |
| REFERENCE STANDARD TEMP | 33.98c   |       |
| REFERENCE STANDARD      | 0.09     | 22:08 |
| BLANK TEST              | 0.00     | 22:09 |
| REFERENCE STANDARD TEMP | 33.97c   |       |
| REFERENCE STANDARD      | 0.09     | 22:09 |
| BLANK TEST              | 0.00     | 22:10 |
| REFERENCE STANDARD TEMP | 33.97c   |       |
| REFERENCE STANDARD      | 0.09     | 22:10 |
| BLANK TEST              | 0.00     | 22:11 |



*R* — — — — — *10/25/11*