

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/31/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:51
INTERNAL STANDARD	VERIFIED	02:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:52
BLANK TEST	0.00	02:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:54



565 Rugh / ~~Rug~~ WESO #162 10/31/11