

**SUPERVISOR TEST**




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/09/2011

Operator Name: CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:41
INTERNAL STANDARD	VERIFIED	11:42
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	11:42
BLANK TEST	0.00	11:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:43
BLANK TEST	0.00	11:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	11:44
BLANK TEST	0.00	11:45



*Charles B. Geisler* #248 11/09/11 11:45 Hrs