

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/15/2011

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11140

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:35
INTERNAL STANDARD	VERIFIED	09:36
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	09:36
BLANK TEST	0.00	09:37
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	09:37
BLANK TEST	0.00	09:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:38
BLANK TEST	0.00	09:39



Handwritten signature and notes:
#248 11/15/11 0938