

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/09/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:56
INTERNAL STANDARD	VERIFIED	23:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:57
BLANK TEST	0.00	23:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:58
BLANK TEST	0.00	23:58
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:59
BLANK TEST	0.00	23:59

