

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/25/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:13
INTERNAL STANDARD	VERIFIED	21:13
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	21:14
BLANK TEST	0.00	21:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:15
BLANK TEST	0.00	21:15
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	21:16
BLANK TEST	0.00	21:16

