

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 12/30/2011

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 11190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:58
INTERNAL STANDARD	VERIFIED	02:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:58
BLANK TEST	0.00	02:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:59
BLANK TEST	0.00	03:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:00
BLANK TEST	0.00	03:01



565 Ralph J Bartlett

WCSO #162

12-30-11