

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/18/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:43
INTERNAL STANDARD	VERIFIED	02:43
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:44
BLANK TEST	0.00	02:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:45
BLANK TEST	0.00	02:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:46



SUT Ralph Bartlett WESO #102 1/18/11