

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 01/24/2011

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:38
INTERNAL STANDARD	VERIFIED	02:38
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:41



Sgt Russ Brown wes0#162 1/24/11