

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 02/07/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10270

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:25
INTERNAL STANDARD	VERIFIED	02:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:25
BLANK TEST	0.00	02:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:26
BLANK TEST	0.00	02:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:27
BLANK TEST	0.00	02:28



SGT Ralph Bartlett

WISD #162

02-07-11