

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/15/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10270

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:49
INTERNAL STANDARD	VERIFIED	03:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:49
BLANK TEST	0.00	03:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:50
BLANK TEST	0.00	03:51
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:51
BLANK TEST	0.00	03:52



505 Ring & Brown W.C.S.O. SHERIFF #162 03/15/11