

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 03/30/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:50
INTERNAL STANDARD	VERIFIED	01:50
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:52
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:53
BLANK TEST	0.00	01:53



505 Ralph Bartlett WISO #162 03/30/11