

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/13/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:43
INTERNAL STANDARD	VERIFIED	02:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:43
BLANK TEST	0.00	02:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:44
BLANK TEST	0.00	02:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:45
BLANK TEST	0.00	02:46



SGT *Ralph Bartlett* 144506 #11010 04-13-11