

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/21/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:05
INTERNAL STANDARD	VERIFIED	02:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:07
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	34.05c	
REFERENCE STANDARD	0.09	02:08
BLANK TEST	0.00	02:08



SGT *Ralph Bartlett* WESO #162 04/21/11