

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/26/2011

Operator Name: JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:41
INTERNAL STANDARD	VERIFIED	22:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:41
BLANK TEST	0.00	22:42
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	22:42
BLANK TEST	0.00	22:43
REFERENCE STANDARD TEMP	33.92c	
REFERENCE STANDARD	0.09	22:43
BLANK TEST	0.00	22:44

*J. M. Lopez #99*

