

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/02/2011

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:01
INTERNAL STANDARD	VERIFIED	04:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:01
BLANK TEST	0.00	04:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:02
BLANK TEST	0.00	04:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:03
BLANK TEST	0.00	04:04



SOS Rugby Brewer Wess # 162 5/2/11