

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/11/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:21
INTERNAL STANDARD	VERIFIED	03:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:21
BLANK TEST	0.00	03:22
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:22
BLANK TEST	0.00	03:23
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:23
BLANK TEST	0.00	03:24



505 *Ralph Bartlett* WCSO #162 5/11/11