

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/16/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:12
INTERNAL STANDARD	VERIFIED	02:12
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:12
BLANK TEST	0.00	02:13
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:13
BLANK TEST	0.00	02:14
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:14
BLANK TEST	0.00	02:15



Sgt. Ralph Bartlett W0107162 5-16-11