

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/24/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11010

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:14
INTERNAL STANDARD	VERIFIED	02:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:14
BLANK TEST	0.00	02:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:16
BLANK TEST	0.00	02:16
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:17
BLANK TEST	0.00	02:17



SGS Ralph J Bartlett

WARREN CO SHERIFF #162

05-24-11