

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/02/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:47
INTERNAL STANDARD	VERIFIED	01:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:48
BLANK TEST	0.00	01:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:49
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:50
BLANK TEST	0.00	01:50



SOI *Ralph Bartlett* WESO SHERIFF #16C 06-02-11