

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/16/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:06
INTERNAL STANDARD	VERIFIED	02:06
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:07
BLANK TEST	0.00	02:08
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:08
BLANK TEST	0.00	02:09



SGT *Ralph Bartlett* WESO #162 06-16-11