

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/22/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:37
INTERNAL STANDARD	VERIFIED	02:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:38
BLANK TEST	0.00	02:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:40



SGS *Ralph Bartlett* were checked #162 06-22-11