

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/19/2011

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:09
INTERNAL STANDARD	VERIFIED	03:09
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:09
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:10
BLANK TEST	0.00	03:11
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:11
BLANK TEST	0.00	03:12

