

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dnt: 144506

Date: 07/25/2011

Operator Name: JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 11040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:05
INTERNAL STANDARD	VERIFIED	03:05
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:06
BLANK TEST	0.00	03:06
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	03:07
BLANK TEST	0.00	03:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:08
BLANK TEST	0.00	03:09

*Handwritten signature and number #299*

